



CE Declaration of Conformity

We: Labconco Corporation
8811 Prospect Avenue
Kansas City, MO 64132 USA
Ph. 816-333-8811 Fax. 816-363-0130
labconco@labconco.com

being the representative party solely responsible for this certification, hereby declare that the following products:

Equipment: Laboratory Equipment – Purifier CELL Logic+ Series Biological Safety Cabinets

Type, Model:	3y239xx-21, 31, 41, 51, 61, 71	Type A2 3' Wide, 9" Sash
	3y249xx-21, 31, 41, 51, 61, 71	Type A2 4' Wide, 9" Sash
	3y259xx-21, 31, 41, 51, 61, 71	Type A2 5' Wide, 9" Sash
	3y269xx-21, 31, 41, 51, 61, 71	Type A2 6' Wide, 9" Sash
	3y348xx-21, 31, 41, 51, 61, 71	Type B2 4' Wide, 8" Sash
	3y368xx-21, 31, 41, 51, 61, 71	Type B2 6' Wide, 8" Sash

y = 2, 3, 4 x = any number 0 through 9

NOTE: Above model numbers may be followed by a 5-digit number which signify an approved modification.
This modification does not impact the compliance to standards listed below.

Manufacturer: Labconco Corporation
8811 Prospect Avenue
Kansas City, MO 64132 USA
Ph. 816-333-8811 Fax. 816-363-0130
labconco@labconco.com

are in accordance with the following directives:

2014/30/EU	Electromagnetic Compatibility Directive	2015/863/EU	RoHS3 Directive
2014/35/EU	Low Voltage Directive		
2006/42/EU	Machinery Directive		
2011/65/EU	RoHS Directive		

are in conformity with the standards listed below:

EN61010-1:2010	EN61000-4-3:2020
EN61326-1:2013	EN61000-4-4:2012
EN55011:2016+A1:2017+A11:2020	EN61000-4-5:2014+A1:2017
EN61000-3-2:2014	EN61000-4-6:2014
EN61000-3-3:2013	EN61000-4-11:2020
EN61000-4-2:2009	EN63000:2018

when installed and operated in accordance with the manufacturers installation and operating instructions.

Brandon Gray
Vice President of Engineering
Labconco Corporation
Kansas City, MO USA

29 JAN 2024

Date of issue



UKCA Declaration of Conformity

We: Labconco Corporation
8811 Prospect Avenue
Kansas City, MO 64132 USA
Ph. 816-333-8811 Fax. 816-363-0130
labconco@labconco.com

being the representative party, in accordance with UK government guidance, solely responsible for this certification, hereby declare that the following product:

Equipment: Laboratory Equipment – Purifier CELL Logic+ Series Biological Safety Cabinets

Type, Model:	3y239xx-21, 31, 41, 51, 61, 71	Type A2 3' Wide, 9" Sash
	3y249xx-21, 31, 41, 51, 61, 71	Type A2 4' Wide, 9" Sash
	3y259xx-21, 31, 41, 51, 61, 71	Type A2 5' Wide, 9" Sash
	3y269xx-21, 31, 41, 51, 61, 71	Type A2 6' Wide, 9" Sash
	3y348xx-21, 31, 41, 51, 61, 71	Type B2 4' Wide, 8" Sash
	3y368xx-21, 31, 41, 51, 61, 71	Type B2 6' Wide, 8" Sash

y = 2, 3, 4 x = any number 0 through 9

NOTE: Above model numbers may be followed by a 5-digit number which signify an approved modification. This modification does not impact the compliance to standards listed below.

Manufacturer: Labconco Corporation
8811 Prospect Avenue
Kansas City, MO 64132 USA
Ph. 816-333-8811 Fax. 816-363-0130
labconco@labconco.com

the object of the declaration described above is in conformity with the relevant UK Statutory Instruments (and their amendments):

S.I. 2016/1091	Electromagnetic Compatibility Directive
S.I. 2016/1101	Low Voltage Directive
S.I. 2008/1597	Machinery Directive
S.I. 2012/3032	RoHS Directive

are in conformity with the standards listed below:

EN61010-1:2010	EN61000-4-3:2020
EN61326-1:2013	EN61000-4-4:2012
EN55011:2016+A1:2017+A11:2020	EN61000-4-5:2014+A1:2017
EN61000-3-2:2014	EN61000-4-6:2014
EN61000-3-3:2013	EN61000-4-11:2020
EN61000-4-2:2009	EN63000:2018

when installed and operated in accordance with the manufacturers installation and operating instructions.

Brandon Gray
Vice President of Engineering
Labconco Corporation
Kansas City, MO USA

29 JAN 2024
Date of issue