CE Declaration of Conformity

We: Labconco Corporation
8811 Prospect Avenue
Kansas City, MO 64132 USA
Ph. 816-333-8811 Fax. 816-363-0130
labconco@labconco.com

being the representative party solely responsible for this certification, hereby declare that the following products:

Equipment: Laboratory Equipment – XPert Filtered Balance Enclosures & XVS Ventilation Stations

Type, Model:
39302xx Series – 2' XPert Balance Enclosure
39303xx Series – 3' XPert Balance Enclosure
39304xx Series – 4' XPert Balance Enclosure
39715xx Series – 5' XPert Balance Enclosure
39716xx Series – 6' XPert Balance Enclosure
39718xx Series – 8' XPert Balance Enclosure
39602xx Series – 2' XPert Easy Order Balance Enclosure
39603xx Series – 3' XPert Easy Order Balance Enclosure
39604xx Series – 4' XPert Easy Order Balance Enclosure
48620xx Series – 2' Protector XVS Ventilation Station
48630xx Series – 3' Protector XVS Ventilation Station
48640xx Series – 4' Protector XVS Ventilation Station

Manufacturer:
Labconco Corporation
8811 Prospect Avenue
Kansas City, MO 64132 USA
Ph. 816-333-8811 Fax. 816-363-0130
labconco@labconco.com

are in accordance with the following directives:

2014/30/EU Electromagnetic Compatibility Directive
2014/35/EU Low Voltage Directive
2006/42/EU Machinery Directive
2015/863/EU RoHS3 Directive

are in conformity with the standards listed below:

EN61010-1:2010
EN61326-1:2013
EN55011:2016+A1:2017+A11:2020
EN61000-3-2:2014
EN61000-3-3:2013
EN61000-4-2:2009
EN61000-4-3:2020
EN61000-4-4:2012
EN61000-4-5:2014+A1:2017
EN61000-4-6:2014
EN61000-4-11:2020

when installed and operated in accordance with the manufacturers installation and operating instructions.

Brandon Gray
Vice President of Engineering
Labconco Corporation
Kansas City, MO USA

17 MAY 2021
Date of issue

UKCA Declaration of Conformity

We: Labconco Corporation
8811 Prospect Avenue
Kansas City, MO 64132 USA
Ph. 816-333-8811 Fax. 816-363-0130
labconco@labconco.com

being the representative party, in accordance with UK government guidance, solely responsible for this certification, hereby declare that the following product:

Equipment: Laboratory Equipment – XPert Filtered Balance Enclosures & XVS Ventilation Stations

Type, Model:
39302xx Series – 2' XPert Balance Enclosure
39303xx Series – 3' XPert Balance Enclosure
39304xx Series – 4' XPert Balance Enclosure
39715xx Series – 5' XPert Balance Enclosure
39716xx Series – 6' XPert Balance Enclosure
39718xx Series – 8' XPert Balance Enclosure
39602xx Series – 2' XPert Easy Order Balance Enclosure
39603xx Series – 3' XPert Easy Order Balance Enclosure
39604xx Series – 4' XPert Easy Order Balance Enclosure
48620xx Series – 2' Protector XVS Ventilation Station
48630xx Series – 3' Protector XVS Ventilation Station
48640xx Series – 4' Protector XVS Ventilation Station

Manufacturer: Labconco Corporation
8811 Prospect Avenue
Kansas City, MO 64132 USA
Ph. 816-333-8811 Fax. 816-363-0130
labconco@labconco.com

the object of the declaration described above is in conformity with the relevant UK Statutory Instruments (and their amendments):

S.I. 2016/1091 Electromagnetic Compatibility Directive
S.I. 2016/1101 Low Voltage Directive
S.I. 2008/1597 Machinery Directive
S.I. 2012/3032 RoHS Directive

are in conformity with the standards listed below:

EN61010-1:2010
EN61326-1:2013
EN55011:2016+A1:2017+A11:2020
EN61000-3-2:2014
EN61000-3-3:2013
EN61000-4-2:2009
EN61000-4-3:2020
EN61000-4-4:2012
EN61000-4-5:2014+A1:2017
EN61000-4-6:2014
EN61000-4-11:2020

when installed and operated in accordance with the manufacturers installation and operating instructions.

Brandon Gray
Vice President of Engineering
Labconco Corporation
Kansas City, MO USA

17 MAY 2021
Date of issue

Labconco P/N 3696041, Rev. G, ECO #M906