



# CE Declaration of Conformity

**We:** Labconco Corporation  
8811 Prospect Avenue  
Kansas City, MO 64132 USA  
Ph. 816-333-8811 Fax. 816-363-0130  
labconco@labconco.com

***being the representative party in the EU solely responsible for this certification, hereby declare that the following product:***

Equipment: Laboratory Equipment – Purifier Logic+ Series Biological Safety Cabinets & PuriCare Series Procedure Stations

Type, Model:	30242xx-21, 31, 41, 51, 71	Type A2 4' Wide, 12" Sash
	31241xx-21, 31, 41, 51, 71	Type A2 4' Wide, 10" Sash
	31242xx-21, 31, 41, 51, 71	Type A2 4' Wide, 12" Sash
	30252xx-21, 31, 41, 51, 71	Type A2 5' Wide, 12" Sash
	31252xx-21, 31, 41, 51, 71	Type A2 5' Wide, 12" Sash
	30262xx-21, 31, 41, 51, 71	Type A2 6' Wide, 12" Sash
	31261xx-21, 31, 41, 51, 71	Type A2 6' Wide, 10" Sash
	31262xx-21, 31, 41, 51, 71	Type A2 6' Wide, 12" Sash

Manufacturer: Labconco Corporation  
8811 Prospect Avenue  
Kansas City, MO 64132 USA  
Ph. 816-333-8811 Fax. 816-363-0130  
labconco@labconco.com

***are in accordance with the following directives:***

2014/30/EU	Electromagnetic Compatibility Directive	2015/863/EU	RoHS3 Directive
2014/35/EU	Low Voltage Directive		
2006/42/EU	Machinery Directive		
2011/65/EU	RoHS Directive		

***are in conformity with the standards listed below:***

EN61010-1:2010	EN61000-4-3:2020
EN61326-1:2013	EN61000-4-4:2012
EN55011:2016+A1:2017+A11:2020	EN61000-4-5:2014+A1:2017
EN61000-3-2:2014	EN61000-4-6:2014
EN61000-3-3:2013	EN61000-4-11:2020
EN61000-4-2:2009	EN63000:2018

***when installed and operated in accordance with the manufacturers installation and operating instructions.***

**Brandon Gray**  
Vice President of Engineering  
Labconco Corporation  
Kansas City, MO USA

29 JAN 2024  
Date of issue



# UKCA Declaration of Conformity

**We:** Labconco Corporation  
8811 Prospect Avenue  
Kansas City, MO 64132 USA  
Ph. 816-333-8811 Fax. 816-363-0130  
labconco@labconco.com

***being the representative party, in accordance with UK government guidance, solely responsible for this certification, hereby declare that the following product:***

Equipment: Laboratory Equipment – Purifier Logic+ Series Biological Safety Cabinets & PuriCare Series Procedure Stations

Type, Model:	30242xx-21, 31, 41, 51, 71	Type A2 4' Wide, 12" Sash
	31241xx-21, 31, 41, 51, 71	Type A2 4' Wide, 10" Sash
	31242xx-21, 31, 41, 51, 71	Type A2 4' Wide, 12" Sash
	30252xx-21, 31, 41, 51, 71	Type A2 5' Wide, 12" Sash
	31252xx-21, 31, 41, 51, 71	Type A2 5' Wide, 12" Sash
	30262xx-21, 31, 41, 51, 71	Type A2 6' Wide, 12" Sash
	31261xx-21, 31, 41, 51, 71	Type A2 6' Wide, 10" Sash
	31262xx-21, 31, 41, 51, 71	Type A2 6' Wide, 12" Sash

Manufacturer: Labconco Corporation  
8811 Prospect Avenue  
Kansas City, MO 64132 USA  
Ph. 816-333-8811 Fax. 816-363-0130  
labconco@labconco.com

***the object of the declaration described above is in conformity with the relevant UK Statutory Instruments (and their amendments):***

S.I. 2016/1091	Electromagnetic Compatibility Directive
S.I. 2016/1101	Low Voltage Directive
S.I. 2008/1597	Machinery Directive
S.I. 2012/3032	RoHS Directive

***are in conformity with the standards listed below:***

EN61010-1:2010	EN61000-4-3:2020
EN61326-1:2013	EN61000-4-4:2012
EN55011:2016+A1:2017+A11:2020	EN61000-4-5:2014+A1:2017
EN61000-3-2:2014	EN61000-4-6:2014
EN61000-3-3:2013	EN61000-4-11:2020
EN61000-4-2:2009	EN63000:2018

***when installed and operated in accordance with the manufacturers installation and operating instructions.***

**Brandon Gray**  
Vice President of Engineering  
Labconco Corporation  
Kansas City, MO USA

29 JAN 2024  
Date of issue