



Declaration of conformity

We: Labconco Corporation
8811 Prospect Avenue
Kansas City, MO 64132USA
Ph. +1 816-333-8811 Fax. 816-363-0130
www.labconco.com

being the representative party solely responsible for this certification, hereby declare that the following products:

Equipment: ReVo Class II Microbiological Safety Cabinet
Type, Models: 3R2xxxxxx and 3T2xxxxxx
Manufacturer: LaboGene A/S
Bjarkesvej 5, 3450 Allerød, Denmark
Ph. +45 39 40 25 66
Mail info@labogene.com

are in accordance with the following directives:

2006/42/EC - Machinery Directive, as amended
2006/95/EC - Low Voltage Directive, as amended
2004/108/EC - EMC Directive, as amended
2011/65/EU - RoHS Directive
2012/19/EU - WEEE Directive

are in conformity with the standards listed below:

EN 12469 - Biotechnology - Performance criteria for microbiological safety cabinets
EN ISO 14644 - Cleanrooms and associated controlled environments
EN 61010-1 - Safety requirements for electrical equipment for measurement, control and laboratory use - General requirements
EN 61000-6-3, EN 61000-6-1 - Electromagnetic compatibility - Generic immunity/emission standard
EN ISO 11201 – Acoustics – Noise emitted by machinery and equipment

when installed and operated in accordance with the manufacturer's installation and operating instructions.

Allerød, September 2022

Rasmus Sørensen
QA Manager
LaboGene A/S, Bjarkesvej 5, 3450 Allerød, Denmark



UKCA Declaration of conformity

We: Labconco Corporation
8811 Prospect Avenue
Kansas City, MO 64132USA
Ph. +1 816-333-8811 Fax. 816-363-0130
www.labconco.com

being the representative party, in accordance with UK government guidance, solely responsible for this certification, hereby declare that the following products:

Equipment: ReVo Class II Microbiological Safety Cabinet
Type, Models: 3R2xxxxxx and 3T2xxxxxx
Manufacturer: LaboGene A/S
Bjarkesvej 5, 3450 Allerød, Denmark
Ph. +45 39 40 25 66
Mail info@labogene.com

The object of the declaration described above with the relevant UK Statutory Instruments:

- S.I. 2008/1597 - The Supply of Machinery (Safety) as amended
- S.I. 2016/1101 - The Electrical Equipment (Safety) as amended
- S.I. 2016/1091 - Electromagnetic Compatibility as amended
- S.I. 2012/3032 - Restriction of the Use of Certain Hazardous Substances in Electrical and Electronic Equipment as amended
- S.I. 2013/3113 - Waste Electrical and Electronic Equipment as amended

are in conformity with the standards listed below:

- EN 12469 - Biotechnology - Performance criteria for microbiological safety cabinets
- EN ISO 14644 - Cleanrooms and associated controlled environments
- EN 61010-1 - Safety requirements for electrical equipment for measurement, control and laboratory use - General requirements
- EN 61000-6-3, EN 61000-6-1 - Electromagnetic compatibility - Generic immunity/emission standard
- EN ISO 11201 - Acoustics - Noise emitted by machinery and equipment

when installed and operated in accordance with the manufacturer's installation and operating instructions.

Allerød, September 2022

Rasmus Sørensen
QA Manager
LaboGene A/S, Bjarkesvej 5, 3450 Allerød, Denmark
27092022QA

