


IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE)
CB SCHEME**CB TEST CERTIFICATE**

Product	SKAN pure2 aseptic / aseptic-toxic isolator
Name and address of the applicant	SKAN AG Kreuzstraße 5, 4123 Allschwil Switzerland
Name and address of the manufacturer	SKAN AG Kreuzstraße 5, 4123 Allschwil Switzerland
Name and address of the factory	Deutsche Mechatronics GmbH Friedrich-Wilhelm-Str.14, 53894 Mechernich Germany
<i>Note: When more than one factory, please report on page 2ff.</i>	
Ratings and principal characteristics	220-240V, 1/N/PE AC (north America (US/CA): L1/L2/PE AC), 50/60Hz, max. P total 3,5 kW
Trademark / Brand (if any)	
Customer Test Facility (CTF) Stage used	
Model / Type Ref.	pure2 1800 4 gloves 1200 2 gloves, M1
Additional information (if necessary may also be reported on page 2ff.)	Re-issue DE-6-G5210047-A2/M1 (issued 2024-08-19) due to non-technical change.
A sample of the product was tested and found to be in conformity with	IEC 61010-1:2010 IEC 61010-1:2010/AMD1:2016
	<i>National differences:</i> EU Group Differences, CA, JP, US
As shown in the Test Report Ref. No. which forms part of this Certificate	G0M-2007-9166-SCM001X-V02, G0M-2007-9166-SCM001X-V02-A2

This CB Test Certificate is issued by the National Certification Body



Eurofins Product Service GmbH
Storkower Strasse 38C
DE-15526 Reichenwalde b. Berlin
Germany



Date: 2024-12-12

Signature: Holger Schaffner